

Nitric Oxide (NO) Colorimetric Assay Kit

Cat No: HR3BC1123

For research use only

Overview

| | |
|---------------------|--|
| Detection Method | Colorimetric method |
| Storage | 2-8? |
| Instrument | Microplate reader(540-550 nm) |
| Assay Time | 55 min |
| Validity | 6 |
| Assay Type | Quantitative |
| Sample Type | Serum, plasma, tissue, saliva |
| Synonyms | NO |
| Instrument | Microplate reader(540-550 nm) |
| Detection Principle | NO is easily oxidized to form NO ₂ ⁻ in vivo or in aqueous solution, and a reddish azo compound is formed with the color developing agent, and the concentration of the azo compound is linearly related to the concentration of NO. The concentration of NO can be calculated indirectly by measuring the OD value at 550 nm. |
| Reagents | Normal saline (0.9% NaCl), PBS (0.01 M, pH 7.4) |
| Labware | Micropipettor, Vortex mixer, Centrifuge |
| Size | 96T |
| Sensitivity | 0.16 ?mol/L |
| Detection Range | 0.16-100 ?mol/L |
| Recovery Rate | 102 |